



UNIVERSITÀ
DEGLI STUDI
FIRENZE

FLORE

Repository istituzionale dell'Università degli Studi di Firenze

Efficacy of ProTaper universal retreatment files in removing filling materials during root canal retreatment.

Questa è la Versione finale referata (Post print/Accepted manuscript) della seguente pubblicazione:

Original Citation:

Efficacy of ProTaper universal retreatment files in removing filling materials during root canal retreatment / V. Giuliani ; R. Cocchetti ; G. Pagavino.. - In: JOURNAL OF ENDODONTICS. - ISSN 0099-2399. - STAMPA. - 34:(2008), pp. 1381-1384.

Availability:

This version is available at: 2158/345766 since:

Terms of use:

Open Access

La pubblicazione è resa disponibile sotto le norme e i termini della licenza di deposito, secondo quanto stabilito dalla Policy per l'accesso aperto dell'Università degli Studi di Firenze (<https://www.sba.unifi.it/upload/policy-oa-2016-1.pdf>)

Publisher copyright claim:

(Article begins on next page)

Efficacy of ProTaper Universal Retreatment Files in Removing Filling Materials during Root Canal Retreatment

Valentina Giuliani, DMD, Roberto Cocchetti, MD, and Gabriella Pagavino, MD, DMD

Abstract

The aim of this study was to evaluate the efficacy of the ProTaper Universal System rotary retreatment system and of Profile 0.06 and hand instruments (K-file) in the removal of root filling materials. Forty-two extracted single-rooted anterior teeth were selected. The root canals were enlarged with nickel-titanium (NiTi) rotary files, filled with gutta-percha and sealer, and randomly divided into 3 experimental groups. The filling materials were removed with solvent in conjunction with one of the following devices and techniques: the ProTaper Universal System for retreatment, ProFile 0.06, and hand instruments (K-file). The roots were longitudinally sectioned, and the image of the root surface was photographed. The images were captured in JPEG format; the areas of the remaining filling materials and the time required for removing the gutta-percha and sealer were calculated by using the nonparametric one-way Kruskal-Wallis test and Tukey-Kramer tests, respectively. The group that showed better results for removing filling materials was the ProTaper Universal System for retreatment files, whereas the group of ProFile rotary instruments yielded better root canal cleanliness than the hand instruments, even though there was no statistically significant difference. The ProTaper Universal System for retreatment and ProFile rotary instruments worked significantly faster than the K-file. The ProTaper Universal System for retreatment files left cleaner root canal walls than the K-file hand instruments and the ProFile Rotary instruments, although none of the devices used guaranteed complete removal of the filling materials. The rotary NiTi system proved to be faster than hand instruments in removing root filling materials. (*J Endod* 2008;34:1381–1384)

Key Words

Gutta-percha removal, ProTaper Universal, root canal retreatment, rotary NiTi instruments

From the Department of Endodontics, University of Florence, Florence, Italy.

Address requests for reprints to Valentina Giuliani, Department of Endodontics, University of Florence, Via Ponte di Mezzo 46/48, Florence, Italy. E-mail address: valesoho@alice.it. 0099-2399/\$0 - see front matter

Copyright © 2008 American Association of Endodontists. doi:10.1016/j.joen.2008.08.002

Endodontic failure might occur in case of persistence of bacteria in the root canal system as a consequence of insufficient cleaning, inadequate obturation, or when there is coronal leakage (1). The failure might be successfully remedied by orthograde retreatment or, if that is not possible, by a surgical procedure (2).

Nonsurgical procedures require the complete removal of filling materials from the endodontic space to obtain 3-dimensional cleaning, shaping, and obturation of the root canal system (3).

Many techniques with rotary nickel-titanium (NiTi) instruments (4–7), ultrasonic instruments (8–10), heat pluggers (11, 12), and manual instruments with chemical solvents (chloroform, eucalyptol, orange oil) have been proposed for removing root filling materials (13–15). Rotary Ni-Ti instruments proved to be effective (4, 5, 16) and time-saving (4, 17, 18) in removing filling materials. However, none of the several treatment alternatives seems to guarantee canal walls that are completely free of debris (4, 13, 17).

Progressively tapered Ni-Ti rotary files, ProTaper, were developed in 2001. ProTaper instruments (Dentsply Maillefer, Ballaigues, Switzerland) have a convex triangular cross-sectional design with different shafts (19).

Recently, a new NiTi rotary system, ProTaper Universal Tulsa (Dentsply Tulsa, Tulsa, OK) was introduced (20). With respect to the original kit, the new system was integrated with 3 new ProTaper retreatment files, D1, D2, D3, two new ProTaper finishing files, F4 and F5, and with the ProTaper obturator and gutta-percha points.

The 3 ProTaper Universal System retreatment files (PTUS) are designed to facilitate the removal of filling material. Each file has different lengths, tapers, and apical tip diameters. The D1 PTUS instrument has an active tip to facilitate initial penetration into the filling material; the D1 instrument has a length of 16 mm, a tip of 0.30 mm, and a 0.09% taper. The D2 PTUS instrument for removal of filling material at the level of the middle third of the root has a length of 18 mm, a tip of 0.25 mm, and a 0.08% taper. The D3 PTUS instrument for apical filling removal with a length of 22 mm, a tip of 0.20 mm, and a 0.07% taper is used to reach the working length.

According to our knowledge, the literature to date contains only a few studies investigating the use of PTUS retreatment files (21, 22).

The purpose of the study was to compare the cleanliness of root canal walls after retreatment with ProFile rotary Ni-Ti instruments (Dentsply Maillefer), PTUS retreatment files, and Hedström files in single-rooted human teeth.

Materials and Methods

Forty-two extracted single-rooted anterior teeth were selected. The teeth were cleaned with an ultrasonic scaler and washed with sterile solution. Preoperative mesiodistal and buccolingual radiographs were taken to verify the presence of a single straight canal. The coronal access cavity was opened by using a high-speed carbide bur and water spray. After removal of the pulp tissue, patency was assured with a size 10 K-file (Dentsply Maillefer), and the working length was defined at the apical foramen. To standardize the samples, the tooth crowns were cut to obtain root canals with a working length of 19 mm.

Canal Preparation

All the samples were prepared by a single operator. The root canal was enlarged by using a crown-down technique. ProFile 0.06 Taper instruments (Dentsply Maillefer)

TABLE 1. Remaining Filling Material in the Whole Root Canal (expressed as percentage area) for Each Group

Technique	N	Mean (standard deviation)
ProFile	14	10.19 (2.3)
ProTaper	14	5.20 (2.66)
Manual	14	11.72 (5.01)

were used in a variable tip sequence from 40–30 until the canal was prepared for instrument size 06/30 at the working length. An electric motor powered at 250 rpm and in a 1:16 ratio (Tecnica/ATR Motor, Pistoia, Italy) was used for all NiTi instruments. Copious irrigation with sodium hypochlorite 5% (Nicolor 5; OGNA, Milan, Italy) was used during the shaping. A final rinse with 10% ethylenediaminetetraacetic acid (Tuboliclean; OGNA) for 2 minutes followed by a rinse with sterile water completed the preparation.

Canal Obturation

The root canal was dried with sterile paper points and obturated with gutta-percha and sealer (Pulp Canal Sealer; Kerr, Romulus, MI) by using continuous wave of warm gutta-percha technique. Back-filling was performed by using thermoplasticized gutta-percha applied with an Obtura II (Obtura Corp, Fenton, MO). The access cavity was restored by using the resin bonding technique.

The quality of the root filling was deemed adequate when no voids could be seen on the mesiodistal and buccolingual radiographs. All the teeth were stored at 37°C in 100% humidity environment for 2 weeks to allow complete setting of the sealer.

Retreatment Technique

The teeth were randomly divided into 3 groups of 14 each. The coronal filling was removed to allow access to the entrance of the canal. A drop of solvent (Endosolv E Septodont of Canada, INC, Cambridge, ON, Canada) was placed in the chamber to soften the gutta-percha; a total of 0.5 mL of solvent was used during the retreatment procedure for each tooth. A low-torque control motor (Tecnica/ATR Motor) in the preset torque levels recommended by the manufacturer for each type of instrument was used in the following 3 experimental groups.

In group A, ProFile System sizes 40/06 30/06 (Dentsply Maillefer) were used in a crown-down technique to remove gutta-percha and sealer. The canals were reinstrumented with ProFile sizes 06/40 and 06/35 in the crown-down manner. The final apical diameter of each root canal was 35 mm.

In group B, PTUS instruments D1, D2, and D3 were used for retreatment in the crown-down technique until D3 reached the working length. Each sample was reprepared with ProTaper Universal Rotary Shaping (S1, S2) and Finishing files (F1, F2, F3) according to the manufacturer’s instructions, until F3 reached the working length. The final apical diameter of each root canal was 0.30 mm.

In group C, gutta-percha and sealer were removed by using Hedström file (Dentsply Maillefer) in a crown-down technique. The re-preparation was done by hand with stainless steel K-files (Dentsply Maillefer), with enlargement to size 35 and step-back increments to size 50.

A total volume of 20 mL of 5% sodium hypochlorite was used as an irrigant for each tooth during the canal re-preparation. Each instrument was discarded after use in 5 canals, and a single operator prepared all the samples.

Removal of filling materials was judged complete when the working length was reached, and no more gutta-percha could be seen on the last instrument used; the time in seconds was recorded. All the teeth were grooved buccolingually with a diamond disk and sectioned longitudinally. Both halves of the root canal were photographed (Nikon Coolpix 4500; Nikon, Melville, NY) under a stereomicroscope at 40× magnification. The photographs of the samples obtained were captured as JPEG images. The remaining gutta-percha and sealer on the split root halves were measured by the Image J 1.33u Program (National Institutes of Health, Bethesda, MD). The evaluation of coded specimens was performed by 2 operators blinded to the techniques and the devices used for retreatment. For each specimen, the arithmetical means of the area of the canal and of remaining gutta-percha and sealer (in millimeters), obtained by the 2 operators, were used to measure the percentage of remaining filling materials for all specimens.

Statistical Analysis

The intraclass correlation coefficient (ρ) was calculated to estimate the reliability of the measurements taken by the 2 examiners. The percentage of remaining filling material and the mean time of gutta-percha removal were evaluated for each group. Descriptive statistics were expressed by means and standard deviations. The one-way analysis of variance test, post hoc Tukey-Kramer honestly significant difference test, and 95% confidence intervals were used to identify differences between the groups at the apical, middle, and coronal levels. The significance level was set at $P < .05$.

Results

The value of the intraclass correlation coefficient was very high ($\rho = 0.99$). Means and standard deviations of the percentage of residual filling materials are reported in Table 1. When the percentage of residual filling materials (gutta-percha and sealer) was analyzed in the whole root canal retreatment, there was a statistically significant difference between the PTUS and Profile techniques ($P < .005$). Group B (5.20 ± 2.66) obtained better results than groups A (10.19 ± 2.30) and C (11.72 ± 5.01). No statistically significant difference was found between groups A and C (Table 1).

The descriptive analysis regarding the differences among the 3 techniques in the apical, middle, and coronal portions of the root are presented in Table 2.

In the coronal and middle third of the retreatment technique, group B obtained better results than groups A and C ($P < .005$). At the same levels, in group A significantly less filling material was observed than in group C ($P < .005$).

In the apical portions, significant differences in the cleanliness of the roots were observed between groups B and C ($P < .005$); no significant difference was observed between groups A or B and between groups A and C.

TABLE 2. Comparison between the 3 Experimental Groups for the Remaining Filling Materials

	n	Group A, Mean (standard deviation)	Group B, Mean (standard deviation)	Group C, Mean (standard deviation)
Coronal portion	14	3.00 (1.96)	1.08 (1.23)	3.00 (1.73)
Middle portion	14	3.76 (1.27)	2.22 (1.26)	4.18 (2.01)
Apical portion	14	3.44 (1.74)	1.89 (1.40)	4.58 (2.73)

TABLE 3. Time Needed for Complete Removal of Filling Materials (in seconds)

Group	n	Mean (standard deviation)
A	14	412.4 (61.3)
B	14	385.5 (92.4)
C	14	532.0 (86.6)

The Tukey-Kramer test showed that there was a statistically significant difference between the groups re-treated with Ni-Ti instruments and Hedström files ($P < .005$). The PTUS and the ProFile System required less time to remove the filling material with respect to the hand instruments (Table 3); there was no statistically significant difference between group A and group B.

Discussion

In the presence of endodontic failure, a nonsurgical approach to the root canal system is preferable to a surgical procedure, even if there is no evidence of a statistically significantly better prognosis (23). The literature reports variable success percentages for retreatment ranging from 40%–100% (23); the variability of the outcome in endodontic retreatment is related to different factors: patient age and the types of teeth treated (24), the presence of alterations in the natural course of the root canals (25), the possibility of removing the coronal restorations to access the pulp chamber (2), the techniques used to remove the existing filling materials, and the possibility of repairing pathologic or iatrogenic defects (26). In a recent cohort clinical study (Toronto Study Phase 3–4) of a total of 126 teeth examined at 4–6 years of follow-up, 104 teeth (83%) were classified as healed. Three outcome predictor factors for orthograde retreatment were identified: quality of previous root filling, presence of perforation, and apical periodontitis (27). Compared with a previous, similar design cohort study (28), the use of engine-driven instruments and dental microscopes could be associated with an increase in the percentage of successes in endodontic retreatment procedures (81% versus 83%); further studies would be beneficial.

The most widely used root filling material is still gutta-percha in conjunction with various sealers, even if a thermoplastic synthetic polymer-base root canal filling material, Resilon (Pentron Clinical Technologies, Wallingford, CT), has become available during the past few years.

In the present *in vitro* study 3 different instruments, 1 manual and 2 rotary, were used for endodontic retreatment. All the retreated teeth had relatively straight root canals that were initially enlarged to size 30/06 to match all the samples in the 3 experimental groups.

As reported in previous studies (7, 12, 17, 29), not one of the experimental techniques guarantees complete removal of the filling materials.

Group B showed a significant efficacy in the removal of the filling materials from the entire canal wall surface, especially in the coronal and middle third, with respect to the other 2 groups (ProFile rotary Ni-Ti instruments and Hedström file).

As previously observed by Hülsmann and Bluhm (4), the presence of different results concerning the cleaning ability of Ni-Ti rotary files could depend on the characteristics of the cross-sectional design of the instruments. PTUS retreatment files remove large amounts of gutta-percha in spirals around the instruments, whereas the U-type cross-section files (ProFile) remove the gutta-percha in small increments that do not adhere to the instruments. The negative cutting angles and the absence of radial land might permit a cutting action rather than a planing action. PTUS retreatment files both soften the gutta-percha by rotation and cut it, whereas the Profile with the U-type cross-section design might not always cut the filling materials. On the basis of the results of

the study presented, this different action is particularly evident in the coronal and middle portions of the root.

In the apical portion of the root, the PTUS performed better than the Hedström files but not the ProFile. This might be explained by the small differences between the tapers and diameters of the 2 devices used in the instrumentation of the apical portions of the root canals.

As suggested in previous studies (21, 22), further root canal refining is necessary because of the apical diameter of the D3 PTUS retreatment file (size 20); the last instrument is designed to reach the working length, but it does not permit a complete cleaning action. The ProFile rotary instruments, on the other hand, allow the removal of filling materials without the need for subsequent root canal refining.

In a previous study (21), a different method was used to assess the cleaning of the canal walls. In this study the teeth were split longitudinally, and residual gutta-percha and sealer were measured linearly. This is not necessarily the best or the most precise method, but it minimizes subjectivity with respect to the use of a scoring system based on scales (7, 18, 29). The average score of one experimental group does not always reflect the original data. Moreover, microcomputer tomography (30), microradiographic technique, and transparent teeth methods (18) represent the most valuable techniques for the qualitative and quantitative evaluation of retreatment procedures.

The introduction of NiTi instruments (4) and the use of solvent (18) have been reported to decrease the time required to remove gutta-percha and sealer. A small amount of solvent containing tetrachloroethylene (Endosolv E) was used at the beginning of the retreatment procedure to soften the coronal filling material to improve the penetration of the files and to avoid the formation of a film of gutta-percha on the canal walls as observed when chloroform was used (31).

Moreover, the active tip of the D1 file might facilitate the penetration of the subsequent files (D2 and D3), as opposed to the shaping files (S1–S2) of the original ProTaper System that cannot penetrate the gutta-percha without fracturing the file tip (4). The nonactive tips of D2 and D3 reduce the incidence of ledging, perforation, and stripping during the removal of filling materials, as opposed to another retreatment instrument system (Mtwo-Retreatment; Sweden & Martina, Padova, Italy), which has active tips for all retreatment instruments (Mtwo R25/.05 and Mtwo R15/.05).

The PTUS retreatment files and Profile Rotary files result in the shortest working time. This concurs with previous studies (4, 5, 21, 29), but at the same time it conflicts with other authors who took significantly less time to achieve clean root canals with manual instruments than with rotary files (17, 32). As suggested, a combination of rotary devices for initial quick removal of gutta-percha and hand instruments to refine and complete the cleaning of the canal especially in the apical third of the root (6, 16) represent the better protocol for obtaining clean canal walls during endodontic retreatment.

Conclusion

Under the experimental conditions PTUS retreatment files left significantly cleaner root canal walls than the K-file hand instruments and the Pro-File rotary instruments. There was a statistically significant difference in favor of NiTi System with respect to the Hedström files regarding the time required to remove gutta-percha.

References

1. Sjogren U, Hagglund B, Sundqvist G, Wing K. Factors affecting the long-term results of Endodontic treatment. *J Endod* 1990;16:498–504.
2. Ruddle C. Nonsurgical retreatment. *J Endod* 2004;30:827–45.
3. Stabholz A, Friedman S. Endodontic retreatment: case selection and technique—part 2: treatment planning for retreatment. *J Endod* 1988;14:607–14.

4. Hülsmann M, Bluhm V. Efficacy, cleaning ability and safety of different rotary NiTi instruments in root canal retreatment. *Int Endod J* 2004;37:468–76.
5. Schirrmester JF, Wrbas KT, Meyer KM, Altenburger MJ, Hellwig E. Efficacy of different rotary instruments for gutta-percha removal in root canal retreatment. *J Endod* 2006;32:469–72.
6. Hülsmann M, Stotz S. Efficacy, cleaning ability and safety of different devices for gutta-percha removal in root canal retreatment. *Int Endod J* 1997;30:227–33.
7. Tasdemir T, Yildirim T, Celik D. Comparative study of removal of current endodontic fillings. *J Endod* 2008;34:326–9.
8. Krell KV, Neo J. The use of ultrasonic endodontic instrumentation in the re-treatment of a paste-filled endodontic tooth. *Oral Surg Oral Med Oral Pathol* 1985;60:100–2.
9. Jeng HW, ElDeed ME. Removal of hard paste fillings from the root canal by ultrasonic instrumentation. *J Endod* 1987;13:295–8.
10. Wilcox LR. Endodontic retreatment: ultrasonic and chloroform as a final step in reinstrumentation. *J Endod* 1989;15:125–8.
11. Friedman S, Stabholz A, Tamse A. Endodontic retreatment: case selection and technique—part 3: retreatment techniques. *J Endod* 1990;16:543–9.
12. Friedman S, Mor C. The success of endodontic therapy: healing and functionally. *J Calif Dent Assoc* 2004;32:267–74.
13. Schirrmester JF, Wrbas KT, Schneider FH, Markus J, Altenburger MJ, Hellwig E. Effectiveness of a hand file and three nickel-titanium rotary instruments for removing gutta-percha in curved root canals during retreatment. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 2006;101:542–7.
14. Wilcox L, Krell KV, Madison S, Rittman B. Endodontic retreatment. Evaluation of gutta-percha and sealer removal and canal reinstrumentation. *J Endod* 1987;13:453–7.
15. Friedman S, Moshonov J, Trope M. Efficacy of removing glass-ionomer cement, zinc oxide eugenol, and epoxy resin sealers from retreated root canals. *Oral Surg Oral Med Oral Pathol Oral Radiol and Endod* 1992;73:609–12.
16. Betti LV, Bramante CM, Quantec SC rotary instruments versus hand files for gutta-percha removal in root canal retreatment. *Int Endod J* 2001;34:514–9.
17. Imura N, Kato AS, Hata G, Uemura M, Toda T, Weine F. A comparison of the relative efficacies of four hand and rotary instrumentation techniques during endodontic retreatment. *Int Endod J* 2003;3:361–6.
18. Ferreira JJ, Rhodes JS, Pitt Ford T. The efficacy of gutta-percha removal using ProFiles. *Int Endod J* 2001;34:267–74.
19. Ruddle C. The ProTaper endodontic system: geometries, features and guidelines for use. *Dentistry Today* 2001;20:60–7.
20. West J. Progressive taper technology: rational and clinical technique for the new ProTaper universal system. *Dentistry Today* 2006;64:66–9.
21. Gu LS, Ling JQ, Wei X, Huang XY. Efficacy of ProTaper Universal rotary retreatment system for gutta-percha removal from root canals. *Int Endod J* 2008;41:288–95.
22. Huang X, Ling J, Gu L. Quantitative evaluation of debris extruded apically by using ProTaper Universal Tulsa Rotare System in endodontic retreatment. *J Endod* 2007;33:1102–5.
23. Paik S, Sechrist C, Torabinejad M. Levels of evidence for the outcome of endodontic retreatment. *J Endod* 2004;30:745–50.
24. Imura N, Pinheiro ET, Gomes BP, Zaia AA, Ferraz CC, Souza-Filho FJ. The outcome of endodontic treatment: a retrospective study of 2000 cases performed by a specialist. *J Endod* 2007;33:1278–82.
25. Gorni FGM, Galliani MM. The outcome of endodontic retreatment: 2 yr follow-up. *J Endod* 2004;30:1–4.
26. Farzaneh M, Abitol S, Friedman S. Treatment outcome in endodontics: the Tronto study—phases I and II: orthograde retreatment. *J Endod* 2004;30:627–33.
27. de Chevigny C, Dao TT, Basrani BR, et al. Treatment outcome in endodontics: the Toronto study—phases 3 and 4: orthograde retreatment. *J Endod* 2008;34:131–7.
28. Farzaneh M, Abitol S, Friedman S. Treatment outcome in endodontics: the Toronto study—phases I and II: orthograde retreatment. *J Endod* 2004;30:627–33.
29. Saad AJ, Al-Hadlaq SM, Al-Katheeri NH. Efficacy of two rotary NiTi instruments in the removal of gutta-percha during root canal retreatment. *J Endod* 2007;33:38–41.
30. Rhodes JS, Ford TR, Lynch JA, Liepins PJ, Curtis RV. Micro-computed tomography: a new tool for experimental endodontology. *Int Endod J* 1999;32:165–70.
31. Sae-Lim V, Rajamanickam I, Lim BK, Lee HL. Effectiveness of Profile.04 taper rotary instruments in endodontic retreatment. *J Endod* 2000;26:100–4.
32. Schirrmester JF, Meyer KM, Hermanns P, Altenburger MJ, Wrbas KT. Effectiveness of hand and rotary instrumentation for removing a new synthetic polymer-based root canal obturation material (Epiphany) during retreatment. *Int Endod J* 2006;39:150–6.