

Biosensors 2016

## Nanostructured photoelectrochemical biosensing platform for cancer biomarker detection

Diego Voccia<sup>1</sup>, Francesca Bettazzi<sup>1</sup>, Serena Laschi<sup>1</sup>, Cristina Gellini<sup>1</sup>, Giangaetano Pietraperzia<sup>1</sup>, Luigi Falciola<sup>2</sup>, Valentina Pifferi<sup>2</sup>, Chiara Ingrosso<sup>3</sup>, Tiziana Placido<sup>3</sup>, Roberto Comparelli<sup>3</sup>, M. Lucia Curri<sup>3</sup>, Iliaria Palchetti<sup>1\*</sup>

<sup>1</sup>Dipartimento di Chimica, Università degli Studi di Firenze, Via della Lastruccia 3, 50019, Sesto Fiorentino, (Fi), Italy

<sup>2</sup>Dipartimento di Chimica, Università degli Studi di Milano, via Golgi 19, 20133, Milano, Italy

<sup>3</sup>CNR-IPCF Istituto per i Processi Chimici e Fisici, Sez. Bari, c/o Dip. Chimica Via Orabona 4, 70126 Bari, Italy

---

### Abstract

The innovative photoelectrochemical properties of multifunctional nanomaterials are here investigated for the development of biosensing platforms for rapid and sensitive detection of a class of cancer biomarker candidates, known as microRNAs. Many different transducers have been proposed, so far, for microRNA detection. Recently, with the emergence of novel photoelectrochemically active species and new detection schemes, photoelectrochemistry has received increasing attention. Gold nanostructures have been, here, used to modify TiO<sub>2</sub> electrodes. The surface of the nanostructured platform has been modified by nucleic acid capture probes (CPs). Biotinylated target miRNAs have been recognized by the specific CPs. The biosensing platform has been incubated with streptavidin alkaline phosphatase and exposed to a proper substrate. The product of the enzymatic reaction has been photoelectrochemically monitored. A compact and hand-held analytical device has been developed in order to have a final prototype in line with the concept of point of care testing.

© 2017 The Authors. Published by Elsevier Ltd. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

Peer-review under responsibility of the organizing committee of Biosensors 2016

**Keywords:** TiO<sub>2</sub> nanocrystals; Au nanorods; genosensor

---

\* Corresponding author: Tel.: + 390554573323.

E-mail address: [iliana.palchetti@unifi.it](mailto:iliana.palchetti@unifi.it)

## 1. Introduction

Nowadays, there is a great interest for the photocatalytic and photoelectrochemical properties of nanomaterials, such as semiconductor nanocrystals, nanoparticles of noble metals, carbon nanotubes or graphene heterostructures, and their applications in biosensing. Colloidal gold nanomaterials show chemical-physical size- and shape-dependent properties, high surface-volume ratio, high density of active sites for the immobilization of molecules, which make them promising for advanced applications. In the present report, the use of titania electrodes, modified with colloidal gold nanorods (AuNRs) has been proposed for the photoelectrochemical detection of miRNA. miRNAs are endogenous, evolutionarily conserved, non-coding RNA molecules of about 20 nucleotides in length, functioning as posttranscriptional gene regulators. Individual miRNAs have been found to regulate cell apoptosis and proliferation and modulate chemotherapy response. So, their detection has potential in diagnosis and therapy of cancer [1]. Many different transducers have been proposed, so far, for microRNA detection [2]. Indeed, recently, with the emergence of novel photoelectrochemically active species and new detection schemes, photoelectrochemistry has received increasing attention.

## 2. Development of a nanostructured photoelectrochemical genosensor based on TiO<sub>2</sub> electrodes and AuNRs

In Figure 1 the scheme of the assay is reported. The capture probe was immobilized on the AuNRs with the well-known chemistry of thiols. The capture-probe-modified electrode was then allowed to react with the analyte (the synthetic, biotinylated, target miRNA). Then, the biotinylated hybrid was exposed to the alkaline-phosphatase-streptavidin conjugate and the product of the enzymatic reaction was electrochemically monitored using chronoamperometry. Phosphorylated ascorbic acid was used as substrate of the biocatalytic reaction in order to recognize the hybridization event on the gold modified electrode surface. The measurements were performed at a wavelength of 470 nm.

## 3. Conclusions

Applying this photoelectrochemical assay, a detection limit of 20 fmole of the target miRNA (RSD of 5%) was obtained.




Figure 1: Scheme of the assay

## References

- [1] Palchetti I, Affinity biosensors for tumor-marker analysis, *Bioanalysis* 2014; 6 (24): 3417-3435.
- [2] D. Voccia, F. Bettazzi, E. Fratini, D. Berti, I. Palchetti, Improving impedimetric nucleic acid detection by using enzyme-decorated liposomes and nanostructured screen-printed electrodes, *Anal. And Bioanal. Chem.* (2016), (doi:10. 1007/ s00216-016-9593-x).

## Acknowledgements

Ministero dell'Istruzione, Università e Ricerca (MIUR) in the framework of PRIN 2012 (grant no.20128ZZS2H) for funding.